


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

21 - 47 High Street, Feltham, Middlesex, TW13 4UN, UK

 <p style="text-align: center;"><b>UKAS</b> CALIBRATION</p> <p style="text-align: center;"><b>4068</b></p> <p style="text-align: center;">Accredited to <b>ISO/IEC 17025:2005</b></p>	<b>Q-Lab Corporation</b>	
	<b>Issue No: 002</b> <b>Issue date: 13 December 2007</b>	
	<b>Stone Hill Road</b> <b>Farnworth</b> <b>Bolton</b> <b>BL4 9TP</b>	<b>Contact: Mr T Hindley</b> <b>Tel: +44 (0)1204 861616</b> <b>Fax: +44 (0)1204 861617</b> <b>E-Mail: info.eu@q.lab.com</b> <b>Website: www.q-lab.com</b>
<b>Calibration performed at the above address only</b>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
RADIOMETRY Spectral responsivity	250 to 800 nm	5.8%	For calibration of Q-Lab type CR10 and CR20 radiometers only
END			